

# EU DECLARATION OF CONFORMITY



**Product:** Labmaster LUCIA™ Vet Analyzer

Labmaster Oy

Rauhalinnantie 31, 20780 Kaarina, FINLAND

**REF**

LM127

**We, as the manufacturer of the device, take sole responsibility for and hereby declare that the above mentioned product meets the provisions of the following Directives:**

- |            |   |
|------------|---|
| 2014/35/EU | Directive on the making available on the market of electrical equipment designed for use within certain voltage limits            |
| 2014/30/EU | Directive on the harmonisation of the laws of the Member States relating to electromagnetic compatibility                         |
| 2011/65/EU | Directive on the restriction of the use of certain hazardous substances in electrical and electronic equipment (RoHS)             |
| 2006/42/EC | Directive on machinery  |
| 2014/53/EU | Directive on the harmonisation of the laws of the Member States relating to the making available on the market of radio equipment |

**Conformity route:**

Self-declaration

**Standards Applied:**

- |                            |  |
|----------------------------|--|
| EN ISO 9001:2015           | Quality management systems — Requirements  |
| EN ISO 14971:2019/A11:2021 | Medical devices - Application of risk management to medical devices  |
| EN 61326-2-6:2013          | Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 2-6: Particular requirements - In vitro diagnostic (IVD) medical equipment- Immunity test requirements for equipment intended to be used in professional healthcare facility environment. |

**Authorised Signatory:**



Name

Position

Date

14.11.2022

The technical documentation for the device is available from the address above.